

Affordable Digital Finger Pulse Oximeter

The MD Spiro / Micro Direct **MD300FT** digital finger pulse oximeter offers the features of a conventional oximeter in a compact, user-friendly, portable configuration. The **MD300FT** accurately measures blood oxygen saturation levels (SpO₂) and heart pulse rate (PR) along with a SpO₂ waveform on an easy-to-read organic light-emitting diode (OLED) display.

Simple to use, this unique device delivers near immediate results; just place the unit on your finger and within seconds, the **MD300FT** will display accurate readings of your SpO₂ and heart pulse rate. Ideal for all medical (hospital, clinics, EMS) and homecare related applications where fast measurement of vital signs is required.

Features

- Two colors OLED display; Adjustable brightness
- Displays SpO₂, PR, SpO₂ waveform (pleth or line)
- Displays signal strength
- Six display modes, convenience to overview the data
- Low voltage indicator, Low power consumption
- 2 AAA alkaline or rechargeable batteries
- Automatic power-off design
- Light, compact, attractive makes it convenient to carry

MD Spiro / Micro Direct

803 Webster Street
 Lewiston, ME 04240
 Telephone 800-588-3381
 Fax 207-786-7280
www.mdspiro.com



Specifications

Patient Range

Adult, Pediatric

SpO₂

Range 35 - 99%
 Resolution 1%
 Accuracy 70 - 100%: +/- 2%
 0 - 69%: unspecified
 Method Dual wavelength LED

Heart (Pulse) Rate

Range 30 - 254 bpm
 Resolution 1 bpm
 Accuracy +/- 2 bpm or +/- 2%

Display

Type Color OLED display (Multi-display modes)
 Parameters SpO₂, PR, waveform (pleth or line)
 Information Low battery indicator

Environmental

Operating Temperature 41° - 104° F
 Storage Temperature 14° - 104° F
 Operating Humidity 15% RH - 95% RH
 Storage Humidity 10% RH - 95% RH

Classification

Type of Protection Internally powered equipment
 Safety IEC standard 61601-1-1

Mechanical

Size 2.28" x 1.26" x 1.33"
 Weight 1.76 ounces (including 2 batteries)

Battery

Type 2 AAA 1.5V Alkaline batteries or
 2 AAA 1.5V rechargeable batteries
 Operation Time Minimum 30 hours for normal operation